


PIN	SIGNAL
1	Valve Test Point*
2	Flow Signal Output
3	Valve Close*
4	Valve Open*
5	Power Common
6	-15 VDC
7	+15 VDC
8	Set Point Input*
9	No Connection
10	Optional Input*
11	Signal Common
12	Signal Common
13	No Connection
14	No Connection
15	Chassis Ground

*Controller Only

FI 01

MODELE 2259 C - 05000 S V
 GAMME 5000-ccm³ N°SERIE 28275 SPE
 ALIM. +/- 15 VDC SORTIE 0-5 VDC
 Etalonnage ramené à 0° et 760 Torr
MKS INSTRUMENTS FRANCE S.A. LE BOURG ET



PV 02

Flo 01

PV02 Flo01 FI01